

Automated Continuous Noninvasive Ward Monitoring

Validation of Measurement Systems Is the Real Challenge

Bernd Saugel, M.D., Phillip Hoppe, M.D., Ashish K. Khanna, M.D., F.C.C.P., F.C.C.M.

About half of all adverse events in hospitalized patients occur on the general care ward.¹ However, acute cardiorespiratory events do not occur out of the blue. Up to 60% of patients have at least one or more abnormal vital signs as early as 4 to 6 h before a cardiac arrest.² Early detection of changes in cardiorespiratory physiology therefore is critical for preventative or therapeutic measures to be effective and to eventually improve patient outcomes. Automated continuous noninvasive ward monitoring may be a promising approach for improving surveillance of general care ward patients at risk for cardiorespiratory events. With numerous continuous ward monitoring devices flooding the healthcare market, it becomes crucial to rigorously test their measurement performance and validate them in different clinical settings before proceeding to larger clinical trials (fig. 1).

In this context, Breteler *et al.*³ need to be commended for their clinical study investigating the measurement performance of novel wireless sensors for heart rate and respiratory rate monitoring published in this issue of ANESTHESIOLOGY. They compared heart rate and respiratory rate measured using two wearable patch sensors, a bed-based system, and a patient-worn monitor against current intensive care unit standard bedside monitoring (multiparameter monitor) in 25 high-risk postsurgery patients after admission to a step-down unit.³ The authors used several statistical tests to compare these noninvasive sensors with the bedside monitor used as the reference method. The primary outcome was the absolute agreement of heart rate and respiratory rate measurements between the test methods



“[T]he impact of continuous ward monitoring on patient-centered outcomes needs to be investigated in large interventional clinical trials.”

with the test method on one axis and the reference method on the other. Different risk zones within the error grid indicate the risk for the patient to be harmed by inappropriate treatment due to measurements with the test methods in comparison with the “true” reference measurement. Visual inspection of the error grid and calculation of the absolute and relative number of measurements within the risk zones enable the clinical relevance of measurement differences to be assessed. The risk zones need to be defined differently for each measured physiologic variable and should be based on pathophysiologic rationale or expert consensus. Error grid analysis is established to validate methods for the measurement of blood glucose⁴ and blood pressure⁵ but not for heart rate or respiratory rate. In the absence of established error-grid risk zones for heart rate or respiratory rate, the authors chose to define the risk zones based on the cutoff

and the reference method. By using Bland–Altman analysis accounting for repeated measurements within subjects, the authors assessed the mean of the differences and the limits of agreement. A clinically acceptable agreement was defined as $\pm 10\%$, ± 5 beats/min, or ± 3 breaths/min in comparison with the reference method. Although the mean of the differences (often termed “bias”) was low, the relatively wide limits of agreement indicate that the precision of agreement of the test methods in comparison with the reference method still need to be improved.

In addition, as a secondary endpoint, the clinical relevance of measurement differences was assessed using error-grid analysis that provides information about the consequences of incorrect treatment decisions triggered by measurements with the test method. An error grid is a scatter plot

Image: Adobe Stock.

This editorial accompanies the article on p. 424. This article has a related Infographic on p. 19A.

Accepted for publication November 4, 2019. Published online first on January 7, 2020. From the Department of Anesthesiology, Center of Anesthesiology and Intensive Care Medicine, University Medical Center Hamburg-Eppendorf, Hamburg, Germany (B.S., P.H.); the Outcomes Research Consortium, Cleveland, Ohio (B.S., A.K.K.); the Department of Anesthesiology, Section on Critical Care Medicine, Wake Forest University School of Medicine, Wake Forest Baptist Health, Winston-Salem, North Carolina (A.K.K.); the Wake Forest Center for Biomedical Informatics, Winston-Salem, North Carolina (A.K.K.); and the Critical Illness, Injury and Recovery Research Center, Wake Forest University School of Medicine Winston-Salem, North Carolina (A.K.K.).

Copyright © 2020, the American Society of Anesthesiologists, Inc. All Rights Reserved. Anesthesiology 2020; 132:407–10. DOI: 10.1097/ALN.0000000000003100




Fig. 1. Validation of the measurement performance of ward monitoring devices is crucial. Before their use in interventional outcomes research studies or clinical practice, automated continuous ward monitoring devices need to be meticulously validated regarding their measurement performance.

boundaries of the Modified Early Warning Score.⁶ Because of the huge amount of data (the authors analyzed more than 700h of monitoring), the authors used color coding to illustrate superimposing pairs of measurements.

Furthermore, because the ability to follow changes in vital signs over time is an important feature of ward monitoring devices Breteler *et al.*³ also analyzed the ability of each of the ward monitoring devices to follow changes in vital

signs over time using four quadrant plots and concordance rates using an exclusion zone (clinically irrelevant changes) of 1 beat/min for heart rate and 1 breath/min for respiratory rate. The concordance rate is the ratio (percentage) of measurements assessed by the test method and the reference method that change in the same direction (decrease or increase) to the sum of all changes. To describe the technical performance and signal stability, the authors also quantified

data loss with any of the devices and demonstrated that 90% of episodes with data loss did not exceed 15 min.

This intricate analysis of four continuous vital signs monitoring devices is timely and of substantial importance to both the clinician and the data scientist. Although current practice patterns show that most patients are still being monitored on an intermittent “spot check” basis on hospital wards, there are already data showing that continuous ward monitoring for early detection of changes in vital signs may improve patient outcome.^{7,8} However, before the use of ward monitoring systems in large-scale interventional outcome studies or their implementation across hospital systems, several challenges need to be addressed.^{9,10} Ward monitoring systems should use small wireless sensors, be easy to use, filter artifacts to decrease alarm fatigue, have seamless integration with most electronic patient records along with artificial intelligence platforms, and allow real-time synchronization of vital sign data. Additionally, sensors should be able to record and process different vital signs to minimize the number of different sensors that need to be attached to the patient. However, most importantly, health-care providers need to have confidence in using innovative ward monitoring systems. This confidence can only be built if ward monitoring systems provide vital sign readings with “clinically acceptable” accuracy and precision.

Therefore, meticulously performed validation studies such as this work by Breteler *et al.*³ are of crucial importance for future ward monitoring research and a meaningful implementation in daily clinical practice. However, we are still missing guidelines or at least consensus statements on how validation studies of ward monitoring systems should be performed and analyzed. Such guidelines could include a checklist outlining how to perform and report validation studies—as for example the Preferred Reporting Items for Systematic Reviews and Meta-Analyses (PRISMA)¹¹—and even more challenging *a priori* definitions of “clinically acceptable” measurement performance based on different statistical tests. In the absence of any guidelines on how to investigate and validate innovative ward monitoring systems, authors need to carefully consider which statistical tests to use to analyze measurement performance and data quality and how to interpret the findings. After meticulous validation, the impact of continuous ward monitoring on patient-centered outcomes needs to be investigated in large interventional clinical trials.

In summary, before their use in interventional outcomes research studies or clinical practice, automated continuous ward monitoring devices need to be meticulously validated regarding their measurement performance. A critical appraisal of the measurement performance and of the clinical impact are prerequisites for a sustainable long-term implementation of automated continuous ward monitoring eventually allowing a change in the culture of patient surveillance on the general care ward.

Competing Interests

Dr. Saugel has received honoraria for consulting, honoraria for giving lectures, and refunds of travel expenses

from Edwards Lifesciences Inc. (Irvine, California); honoraria for consulting, institutional restricted research grants, honoraria for giving lectures, and refunds of travel expenses from Pulsion Medical Systems SE (Munich, Germany); institutional restricted research grants, honoraria for giving lectures, and refunds of travel expenses from CNSystems Medizintechnik (Graz, Austria); institutional restricted research grants from Retia Medical (Valhalla, New York); honoraria for giving lectures from Philips Medizin Systeme Böblingen (Böblingen, Germany); and honoraria for consulting, institutional restricted research grants, and refunds of travel expenses from Tensys Medical Inc. (San Diego, California). Dr. Khanna collaborates with Medtronic (Dublin, Ireland) as a member of the executive advisory board on respiratory monitoring and receives honoraria for these services, which include giving lectures and refunds of travel expenses. Dr. Khanna also serves on the clinical advisory boards for Retia Medical and Linshom Medical (Ellicott City, Maryland) and serves as a consultant for Phillips North America (Andover, Massachusetts), Zoll Medical (Chelmsford, Massachusetts), and Edwards Lifesciences Inc., and as a subject matter expert for the development of the Anesthesia SimStat system for CAE Healthcare (Sarasota, Florida). Dr. Hoppe declares no competing interests.

Correspondence

Address correspondence to Dr. Saugel: bernd.saugel@gmx.de


References

- Andersen LW, Berg KM, Chase M, Cocchi MN, Massaro J, Donnino MW; American Heart Association's Get With The Guidelines(®)-Resuscitation Investigators: Acute respiratory compromise on inpatient wards in the United States: Incidence, outcomes, and factors associated with in-hospital mortality. *Resuscitation* 2016; 105:123–9
- Andersen LW, Kim WY, Chase M, Berg KM, Mortensen SJ, Moskowitz A, Novack V, Cocchi MN, Donnino MW; American Heart Association's Get With the Guidelines(®)-Resuscitation Investigators: The prevalence and significance of abnormal vital signs prior to in-hospital cardiac arrest. *Resuscitation* 2016; 98:112–7
- Breteler MJM, KleinJan EJ, Dohmen DAJ, Leenen LPH, van Hillegersberg R, Ruurda JP, van Loon K, Blokhuis TJ, Kalkman CJ: Vital signs monitoring with wearable sensors in high-risk surgical patients: A clinical validation study. *ANESTHESIOLOGY* 2020; 132:424–39
- Klonoff DC, Lias C, Vigersky R, Clarke W, Parkes JL, Sacks DB, Kirkman MS, Kovatchev B; Error Grid Panel: The surveillance error grid. *J Diabetes Sci Technol* 2014; 8:658–72

5. Saugel B, Grothe O, Nicklas JY: Error grid analysis for arterial pressure method comparison studies. *Anesth Analg* 2018; 126:1177–85
6. Subbe CP, Kruger M, Rutherford P, Gemmel L: Validation of a Modified Early Warning Score in medical admissions. *QJM* 2001; 94:521–6
7. Taenzer AH, Pyke JB, McGrath SP, Blike GT: Impact of pulse oximetry surveillance on rescue events and intensive care unit transfers: A before-and-after concurrence study. *ANESTHESIOLOGY* 2010; 112:282–7
8. Brown H, Terrence J, Vasquez P, Bates DW, Zimlichman E: Continuous monitoring in an inpatient medical–surgical unit: A controlled clinical trial. *Am J Med* 2014; 127:226–32
9. Sessler DI, Saugel B: Beyond “failure to rescue”: The time has come for continuous ward monitoring. *Br J Anaesth* 2019; 122:304–6
10. Khanna AK, Hoppe P, Saugel B: Automated continuous noninvasive ward monitoring: Future directions and challenges. *Crit Care* 2019; 23:194
11. Liberati A, Altman DG, Tetzlaff J, Mulrow C, Gotzsche PC, Ioannidis JP, Clarke M, Devereaux PJ, Kleijnen J, Moher D: The PRISMA statement for reporting systematic reviews and meta-analyses of studies that evaluate health care interventions: Explanation and elaboration. *Ann Intern Med* 2009; 151:W65–94

ANESTHESIOLOGY REFLECTIONS FROM THE WOOD LIBRARY-MUSEUM

Exploding onto Lancaster's Anesthetic Landscape: Dr. Michael A. Becker's Vitalized Air



After earning his D.D.S. from Cincinnati's Ohio College of Dental Surgery, Michael Augustus Becker (1866 to 1938) married in his home state before moving his new wife to Lancaster, Pennsylvania. As manager of the city's "Albany Dentists" franchise (its trade card obverse with inserts from reverse, *above*), he was in charge of generating "Vitalized Air" for anesthesia and of vulcanizing rubber for dentures. In the latter capacity, Dr. Becker stood next to a pressurized vulcanizer "when it suddenly burst." Lancaster's *Morning News* reported that boiling water and steam severely burned Becker's face, and he was "unable to use his eyes for some time after the accident." Indeed, Becker was fortunate NOT to have had: (1) the exploding vulcanizer's shrapnel kill or permanently blind him, (2) an ignition of the Vitalized Air's supplemental mixture (alcohol-chloroform) burn him, or (3) a chain reaction engulf him with explosions involving the office's gasometer or compressed gas cylinder(s). Because nitrous oxide supports combustion, Becker, though temporarily blinded, was indeed a lucky man. In future years, he would advertise his anesthetic mixture as "Becker's Vitalized Air." (Copyright © the American Society of Anesthesiologists' Wood Library-Museum of Anesthesiology.)

Melissa L. Coleman, M.D., Penn State College of Medicine, Hershey, Pennsylvania, and George S. Bause, M.D., M.P.H., Case Western Reserve University, Cleveland, Ohio.