

Fig. 1. A new and inexpensive device for delivering a topical local anesthetic.

Anesthesiology 56:155-156, 1982

Hazard of Separate Low and High Flow O2 Flowmeters: An Interim Solution

To the Editor:—The publication of the American National Standard on Minimum Performance and Safety Requirements for Components and Systems of Continuous Flow Anesthesia Machines for Human Use Z79.8.1979* drew attention once again to the hazards in present day anesthesia apparatus. The successful legal suits against hospitals for failing to remove from use apparatus incorporating documented hazards led to the recommendation that hospitals update their apparatus without delay.¹

Mazze² reported the anoxic death of a healthy 45year-old man scheduled for hemorrhoidectomy. The anesthesia machine was equipped with two separate O_2 flowmeters, one calibrated in milliliters, the other in liters. The wrong O_2 flowmeter was used, resulting in 4 l of N_2O being delivered with 200 ml of O_2 , rather than 2 l of O_2 .

Unfortunately many hospitals have many apparatus incorporating this hazard. Even though one large self-insured hospital organization estimated it would be less expensive to replace 200–300 obsolete gas machines than to settle another suit for anoxic brain damage, the individual hospital may have budget difficulty in replacing all such apparatus immediately.

As a temporary measure pending their phased replacement we have fitted guards to the low flow O_2 and N_2O controls on our $Ohio^{\oplus}$ 2000 model apparatus. The

^{*} American National Standards Institute, 1430 Broadway, New York, New York 10018.




Fig. 1. Guards attached to the low flow controls for N_2O and O_2 to alert the user to the flowmeter being used.

guards do not prevent the use of these controls, but the user receives a tactile reminder each time the low flow control is adjusted (fig. 1). The hospital's Engineering Department made the guards from 1-mm thick stainless steel and pop rivetted them onto the frame of the plastic flowmeter shield. A similar pattern could be made for the Ohio® DM 5000 or other apparatus where it may be needed.

LESLIE RENDELL-BAKER, M.D. ORVILLE L. KLEIN, A.R.T. PREMILA CHARLES, M.D.

Department of Anesthesiology Loma Linda University Medical Center Loma Linda, California 92357

REFERENCES

- Rendell-Baker L: Update anesthesia equipment to reduce risks. Hospitals 54:49-52, 1980
- Mazze RI: Therapeutic misadventures with oxygen delivery systems: the need for continuous in-line oxygen monitors. Anesth Analg (Cleve) 51:787-792, 1972

(Accepted for publication August 18, 1981.)

Anesthesiology 56:156, 1982

Cardiopulmonary Resuscitation of Late-Pregnant Women

To the Editor:—Cardiac arrest may result from the accidental intravascular injection of more than the test-dose of 0.75 per cent bupivacaine during attempted extradural block. I have detailed information on five such complications occurring in healthy gravidae scheduled for cesarean section. Characteristically, there was a short grand mal seizure followed by disappearance of pulse and blood pressure. All five women required external cardiac compression for more than 10 minutes. In three, the infant was delivered with dispatch, and the mothers survived, two with no after-effects but amnesia. In the other two gravidae, delivery of the fetus was delayed for several minutes; both mothers suffered irreversible brain damage.

Cardiac compression is accomplished most efficiently with the patient supine on a hard surface. In late-pregnant women, however, this position is associated with aortocaval compression and resultant decrease in venous return to the heart. Therefore, delivery of the infant will improve maternal circulation promptly and significantly. If immediate delivery cannot be undertaken safely, expert uterine displacement must be undertaken without delay; manual displacement (to the left and slightly cephalad) is most rational in such a situation.

Prompt delivery of the infant should be considered part of cardiopulmonary resuscitation of late-pregnant women.

GERTIE F. MARX, M.D.

Professor of Anesthesiology

Albert Einstein College of Medicine

Bronx, New York 10461

(Accepted for publication on August 18, 1981.)